

Medi Sense

Model: OCI-16G Expiry Date: 2027.09.18
Size: 25 mm Manufacture Date: 2024.09.19
Type: Rotated wing Lot Number: 240946



Sterile Arteriovenous Fistula Needle Sets



- [Usage]
1. Disinfect the puncture site according to the operation procedures of dialysis treatment.
 2. Confirm the package is not damaged before use, take out the product from the package and remove the needle protector.
 3. Hold the needle handle and puncture the blood vessel at the disinfected site.
 4. After the blood flows into the catheter, close the clamp.
 5. Fix the needle with tape.
 6. Puncture the second needle according to the above steps.
 7. Unscrew the connector protector, connect it with the hemodialysis bloodline, open the clamp, start the blood circulation.
 8. After the treatment, close the clamp, remove the tape, pull out the needle, and dispose it according to the regulations of medical waste.

- [Precautions]
- (1) This product is compatible with other existing brands and models of hemodialysis bloodline on the market at present.
 - (2) This product is sterilized by ethylene oxide, it is sterile, pyrogen free and valid for 3 years.
 - (3) Before using this product, please check the package and the product integrity. It is forbidden to use if the package or the product is damaged.
 - (4) The operation of the product should only be limited to trained medical personnel.
 - (5) After use, all the blood in the tube should be back flow to the patient, then all clamps shall be closed. Finally, remove the Arteriovenous Fistula Needle Sets and other accessories. All of them shall be placed in a special waste container and destroyed according to the clinical regulations.
 - (6) Don't store in high temperature, direct sunlight and humid environment.
 - (7) The duration of device life use is 4h.
 - (8) Read the operation instruction carefully before use.
 - (9) When product breakage or leakage is detected, discontinue use and replace the product or cease treatment in accordance with clinical operation requirements.
 - (10) Some components directly contact with blood are made of PVC (with DEHP). Clinical personnel should pay attention to the possible toxicity to high-risk population (prepubescent males, pregnant and lactating females), and choose substitutes as much as possible. Moreover, the product is not available for infusion of liposoluble liquid and medicine.

The product should be stored at 0-40°C in a dry, ventilated, clean room without corrosive gases.

Chengdu OCI Medical Devices Co., Ltd.
Address: No.2401, West Port Avenue, Southwest Airport Economic Development Zone, Shuangliu District, Chengdu, Sichuan Province, China.
Tel: +86-28-67085880 Fax: +86-28-67085880

EU REP Lepu Medical (Europe) Cooperatief U.A.
Abe Lenstra Boulevard 36, 8448 JB, Heerenvveen, The Netherlands
Tel: +31-515-573399 Fax: +31-515-76

Importer: Superior Healthcare International House # 132 Street #
08 Al Noor Garden Madina Twon Faisalabad
ELI-01389 Registration No. MDIR-EN00000812 M.R.P: Rs.500

Medi Sense

Model: OCI-16G Expiry Date: 2027.09.18
Size: 32 mm Manufacture Date: 2024.09.19
Type: Rotated wing Lot Number: 240946



Sterile Arteriovenous Fistula Needle Sets



- [Usage]
1. Disinfect the puncture site according to the operation procedures of dialysis treatment.
 2. Confirm the package is not damaged before use, take out the product from the package and remove the needle protector.
 3. Hold the needle handle and puncture the blood vessel at the disinfected site.
 4. After the blood flows into the catheter, close the clamp.
 5. Fix the needle with tape.
 6. Puncture the second needle according to the above steps.
 7. Unscrew the connector protector, connect it with the hemodialysis bloodline, open the clamp, start the blood circulation.
 8. After the treatment, close the clamp, remove the tape, pull out the needle, and dispose it according to the regulations of medical waste.

- [Precautions]
- (1) This product is compatible with other existing brands and models of hemodialysis bloodline on the market at present.
 - (2) This product is sterilized by ethylene oxide, it is sterile, pyrogen free and valid for 3 years.
 - (3) Before using this product, please check the package and the product integrity. It is forbidden to use if the package or the product is damaged.
 - (4) The operation of the product should only be limited to trained medical personnel.
 - (5) After use, all the blood in the tube should be back flow to the patient, then all clamps shall be closed. Finally, remove the Arteriovenous Fistula Needle Sets and other accessories. All of them shall be placed in a special waste container and destroyed according to the clinical regulations.
 - (6) Don't store in high temperature, direct sunlight and humid environment.
 - (7) The duration of device life use is 4h.
 - (8) Read the operation instruction carefully before use.
 - (9) When product breakage or leakage is detected, discontinue use and replace the product or cease treatment in accordance with clinical operation requirements.
 - (10) Some components directly contact with blood are made of PVC (with DEHP). Clinical personnel should pay attention to the possible toxicity to high-risk population (prepubescent males, pregnant and lactating females), and choose substitutes as much as possible. Moreover, the product is not available for infusion of liposoluble liquid and medicine.

The product should be stored at 0-40°C in a dry, ventilated, clean room without corrosive gases.

Chengdu OCI Medical Devices Co., Ltd.
Address: No.2401, West Port Avenue, Southwest Airport Economic Development Zone, Shuangliu District, Chengdu, Sichuan Province, China.
Tel: +86-28-67085880 Fax: +86-28-67085880

EU REP Lepu Medical (Europe) Cooperatief U.A.
Abe Lenstra Boulevard 36, 8448 JB, Heerenvveen, The Netherlands
Tel: +31-515-573399 Fax: +31-515-76

Importer: Superior Healthcare International House # 132 Street #
08 Al Noor Garden Madina Twon Faisalabad
ELI-01389 Registration No. MDIR-EN00000812 M.R.P: Rs.500

Medi Sense

Model: OCI-17G Expiry Date: 2027.09.18
Size: 25 mm Manufacture Date: 2024.09.19
Type: Rotated wing Lot Number: 240946



Sterile Arteriovenous Fistula Needle Sets



- [Usage]
1. Disinfect the puncture site according to the operation procedures of dialysis treatment.
 2. Confirm the package is not damaged before use, take out the product from the package and remove the needle protector.
 3. Hold the needle handle and puncture the blood vessel at the disinfected site.
 4. After the blood flows into the catheter, close the clamp.
 5. Fix the needle with tape.
 6. Puncture the second needle according to the above steps.
 7. Unscrew the connector protector, connect it with the hemodialysis bloodline, open the clamp, start the blood circulation.
 8. After the treatment, close the clamp, remove the tape, pull out the needle, and dispose it according to the regulations of medical waste.

- [Precautions]
- (1) This product is compatible with other existing brands and models of hemodialysis bloodline on the market at present.
 - (2) This product is sterilized by ethylene oxide, it is sterile, pyrogen free and valid for 3 years.
 - (3) Before using this product, please check the package and the product integrity. It is forbidden to use if the package or the product is damaged.
 - (4) The operation of the product should only be limited to trained medical personnel.
 - (5) After use, all the blood in the tube should be back flow to the patient, then all clamps shall be closed. Finally, remove the Arteriovenous Fistula Needle Sets and other accessories. All of them shall be placed in a special waste container and destroyed according to the clinical regulations.
 - (6) Don't store in high temperature, direct sunlight and humid environment.
 - (7) The duration of device life use is 4h.
 - (8) Read the operation instruction carefully before use.
 - (9) When product breakage or leakage is detected, discontinue use and replace the product or cease treatment in accordance with clinical operation requirements.
 - (10) Some components directly contact with blood are made of PVC (with DEHP). Clinical personnel should pay attention to the possible toxicity to high-risk population (prepubescent males, pregnant and lactating females), and choose substitutes as much as possible. Moreover, the product is not available for infusion of liposoluble liquid and medicine.

The product should be stored at 0-40°C in a dry, ventilated, clean room without corrosive gases.

Chengdu OCI Medical Devices Co., Ltd.
Address: No.2401, West Port Avenue, Southwest Airport Economic Development Zone, Shuangliu District, Chengdu, Sichuan Province, China.
Tel: +86-28-67085880 Fax: +86-28-67085880

EU REP Lepu Medical (Europe) Cooperatief U.A.
Abe Lenstra Boulevard 36, 8448 JB, Heerenvveen, The Netherlands
Tel: +31-515-573399 Fax: +31-515-76

Importer: Superior Healthcare International House # 132 Street #
08 Al Noor Garden Madina Twon Faisalabad
ELI-01389 Registration No. MDIR-EN00000812 M.R.P: Rs.500

Medi Sense

Model: OCI-17G Expiry Date: 2027.09.18
Size: 32 mm Manufacture Date: 2024.09.19
Type: Rotated wing Lot Number: 240946



Sterile Arteriovenous Fistula Needle Sets



- [Usage]
1. Disinfect the puncture site according to the operation procedures of dialysis treatment.
 2. Confirm the package is not damaged before use, take out the product from the package and remove the needle protector.
 3. Hold the needle handle and puncture the blood vessel at the disinfected site.
 4. After the blood flows into the catheter, close the clamp.
 5. Fix the needle with tape.
 6. Puncture the second needle according to the above steps.
 7. Unscrew the connector protector, connect it with the hemodialysis bloodline, open the clamp, start the blood circulation.
 8. After the treatment, close the clamp, remove the tape, pull out the needle, and dispose it according to the regulations of medical waste.

- [Precautions]
- (1) This product is compatible with other existing brands and models of hemodialysis bloodline on the market at present.
 - (2) This product is sterilized by ethylene oxide, it is sterile, pyrogen free and valid for 3 years.
 - (3) Before using this product, please check the package and the product integrity. It is forbidden to use if the package or the product is damaged.
 - (4) The operation of the product should only be limited to trained medical personnel.
 - (5) After use, all the blood in the tube should be back flow to the patient, then all clamps shall be closed. Finally, remove the Arteriovenous Fistula Needle Sets and other accessories. All of them shall be placed in a special waste container and destroyed according to the clinical regulations.
 - (6) Don't store in high temperature, direct sunlight and humid environment.
 - (7) The duration of device life use is 4h.
 - (8) Read the operation instruction carefully before use.
 - (9) When product breakage or leakage is detected, discontinue use and replace the product or cease treatment in accordance with clinical operation requirements.
 - (10) Some components directly contact with blood are made of PVC (with DEHP). Clinical personnel should pay attention to the possible toxicity to high-risk population (prepubescent males, pregnant and lactating females), and choose substitutes as much as possible. Moreover, the product is not available for infusion of liposoluble liquid and medicine.

The product should be stored at 0-40°C in a dry, ventilated, clean room without corrosive gases.

Chengdu OCI Medical Devices Co., Ltd.
Address: No.2401, West Port Avenue, Southwest Airport Economic Development Zone, Shuangliu District, Chengdu, Sichuan Province, China.
Tel: +86-28-67085880 Fax: +86-28-67085880

EU REP Lepu Medical (Europe) Cooperatief U.A.
Abe Lenstra Boulevard 36, 8448 JB, Heerenvveen, The Netherlands
Tel: +31-515-573399 Fax: +31-515-76

Importer: Superior Healthcare International House # 132 Street #
08 Al Noor Garden Madina Twon Faisalabad
ELI-01389 Registration No. MDIR-EN00000812 M.R.P: Rs.500